


Foot switch

As an alternative to using the Start/Stop buttons, ultrasound generators can also be operated using the TS 8 foot switch.

Description	TS 8
Order No.	531
Figure	
Length connecting cable [m]	3
Dimensions tread surface [mm]	90 × 65
Material of tread surface	Slip-free rubber
Use with HD...	2070.2/2200.2 3100/3200/3400 4050/4100/4200/4400

Mounting

- Connect the plug of the TS 8 to the back side of the ultrasound generator using the remote control connector.